

Weekly Meeting

20/08/2019

ByungChan Lee


Gain Calibration

PMT	HV	MPV
PL_IBS036	1469	704.2 ± 1.9
PR_IBS037	1534	691.7 ± 1.9
NL_IBS020	1405	702.9 ± 2.5
NR_IBS042	1552	701.7 ± 2.5

Gain Calibration : Bar P, N





Fitting : Landau + exponential

Gain Calibration



Gain Calibration

BAR F / GAIN DECREASING



2019 – 08 – 13


MPV : L 730 / R 705



2019 – 08 – 16

MPV : L 495 / R 511

Gain Calibration



- Left PMT & Bakelite was not firmly attached to the bar
- Need to reassemble : 08/21 (tomorrow)